

Test program Separation degrees EN ISO 11143:2002-12
Dental Equipment-Amalgam Separators

Test Date: 10. - 14.02.2006

Expert: Hlg

Applicant:

confidential

UK

Test object: Surf
Classification: Type 4: Filtration and Sedimentation system
Maximum water flow rate: 0,5 l/min

Amalgam sample: Becker Meßtechnik GmbH, 71364 Winnenden, Germany
ISO 11143 Standardprobe Charge : 100416-11/04
Membrane filters: Schleicher + Schuell AE100 + ME29 + ME28
Balance: Kern 434 Ser. No.: 118100

Number of tests performed: 6

Separation degrees of empty amalgam separator with maximum flow rate:

No.		
	1	96,2 %
	2	96,8 %
	3	96,0 %
Average:		96,3 %

Separation degrees of full amalgam separator with maximum flow rate:

No.		
	1	96,7 %
	2	96,2 %
	3	96,6 %
Average:		96,5 %

Value of the efficiency with maximum flow rate: **96,3%**

The R.C.H. Dentserv LTD Amalgam Separator Surf pass the separation degrees test.

Essen, 14.02.2006

For the contents:


Dipl. Ina. Jan Hilaers